
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000694
8:29 16/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:12
STERILIZING 3:12
136.0 C 223kPa 3:12
136.4 C 231kPa 3:42
136.5 C 233kPa 4:12
136.4 C 229kPa 4:42
136.4 C 233kPa 5:12
136.3 C 232kPa 5:42
136.4 C 233kPa 6:12
136.4 C 233kPa 6:42
136.4 C 233kPa 6:42
VENTING 6:42
AIR DRYING 7:35
CYCLE COMPLETE 7:38

Digital Signature #
F884A3E7C543002E
