
Midmark M11 - v1.0.5
Total Cycles: 1226

Name: RITTER2
Sterilizer ID: V2313228
06/27/2022 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.1 F 0.0
2:00 156.1 F 2.7
4:00 168.1 F 6.3
6:00 190.7 F 11.0
8:00 213.5 F 5.0
10:00 227.3 F 10.2
12:00 229.9 F 7.8
14:00 242.0 F 13.1
16:00 241.8 F 11.9
18:00 254.5 F 18.5
20:00 265.8 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.1 F 30.0
2:00 272.9 F 30.1
3:00 272.2 F 29.7
4:00 271.6 F 29.4

Min 271.1 F 29.1
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 214.1 F 0.0
10:00 222.1 F 0.0
15:00 220.7 F 0.0
20:00 218.7 F 0.0
25:00 219.1 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

06/27/2022 10:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____