
Midmark M11 - v1.0.0
Total Cycles: 2630

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2022 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.5 F 0.0
2:00 199.4 F 5.4
4:00 209.9 F 10.6
6:00 222.5 F 5.9
8:00 234.4 F 11.8
10:00 235.5 F 9.5
12:00 238.5 F 9.0
14:00 248.9 F 15.2
16:00 261.5 F 22.3
18:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.3 F 29.2
2:00 271.9 F 29.4
3:00 272.1 F 29.1
4:00 271.8 F 29.3

Min 271.5 F 28.8
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 235.6 F 0.0
10:00 236.0 F 0.0
15:00 228.9 F 0.0
20:00 226.0 F 0.0
25:00 227.6 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:14
CYCLE COMPLETE

11/30/2022 03:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____