
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000108
9:41 27/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:53
STERILIZING 3:12
136.4 C 224kPa 3:12
136.4 C 222kPa 3:42
136.3 C 224kPa 4:12
136.3 C 223kPa 4:42
136.4 C 223kPa 5:12
136.4 C 224kPa 5:42
136.3 C 224kPa 6:12
136.2 C 221kPa 6:42
136.2 C 221kPa 6:42
VENTING 6:42
AIR DRYING 7:45
CYCLE COMPLETE 9:55

Digital Signature #
249CF675F8334358
