
Midmark M11 - v1.0.0
Total Cycles: 1275

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2021 06:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 73.5 F 0.0
4:00 111.8 F 0.0
6:00 109.3 F 0.0
8:00 141.8 F 0.1
10:00 172.3 F 0.2
12:00 196.5 F 0.3
14:00 212.0 F 3.1
16:00 222.1 F 5.5
18:00 229.3 F 8.4
20:00 237.4 F 11.8
22:00 232.6 F 7.8
24:00 244.6 F 13.1
26:00 242.3 F 11.3
28:00 239.5 F 10.0
30:00 252.6 F 16.3
32:00 263.3 F 22.5
34:00 272.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.7 F 28.7
2:00 271.7 F 27.9
3:00 272.1 F 28.5
4:00 272.4 F 28.4

Min 271.3 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 212.4 F 0.0
10:00 208.1 F 0.0
15:00 206.0 F 0.0
20:00 206.3 F 0.0
25:00 202.7 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 34:18
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:12:35
CYCLE COMPLETE

10/19/2021 07:53 AM
ACCEPTED: X
REJECTED:
Operator: U01