
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000416
7:57 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:51
STERILIZING 3:57
136.0 C 228kPa 3:57
136.4 C 233kPa 4:27
136.4 C 233kPa 4:57
136.3 C 231kPa 5:27
136.4 C 233kPa 5:57
136.1 C 229kPa 6:27
136.0 C 229kPa 6:57
136.2 C 231kPa 7:27
136.2 C 231kPa 7:27
VENTING 7:27
AIR DRYING 8:19
CYCLE COMPLETE 8:28

Digital Signature #
BE736597716807DE
