
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001831
8:27 12/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:49
STERILIZING 3:46
136.0 C 220kPa 3:46
135.9 C 221kPa 4:16
136.1 C 223kPa 4:46
136.3 C 225kPa 5:16
136.2 C 223kPa 5:46
136.3 C 224kPa 6:16
136.2 C 223kPa 6:46
136.2 C 223kPa 7:16
136.2 C 223kPa 7:16
VENTING 7:16
AIR DRYING 8:06
CYCLE COMPLETE 38:06

Digital Signature #
2C220A925F429CCD
