

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2259

Name: RITTER2  
Sterilizer ID: V2313228  
04/18/2023 08:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.3 F 0.0  
2:00 141.7 F 1.2  
4:00 147.8 F 3.9  
6:00 175.1 F 8.0  
8:00 214.9 F 12.8  
10:00 223.5 F 7.4  
12:00 228.0 F 12.6  
14:00 232.6 F 9.7  
16:00 235.9 F 9.1  
18:00 246.5 F 14.3  
20:00 258.6 F 20.8  
22:00 268.7 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 30.1  
1:00 272.1 F 30.3  
2:00 272.3 F 30.6  
3:00 272.4 F 30.6  
4:00 272.2 F 30.6

Min 271.7 F 30.0  
Max 272.8 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 221.0 F 0.0  
10:00 236.9 F 0.0  
15:00 222.5 F 0.0  
20:00 219.8 F 0.0  
25:00 219.1 F 0.0  
30:00 219.5 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 23:41  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:01:48  
CYCLE COMPLETE

04/18/2023 09:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_