

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1048

Name: RITTER2  
Sterilizer ID: V2313228  
04/28/2022 01:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.7 F 0.0  
2:00 173.7 F 3.4  
4:00 186.1 F 8.1  
6:00 208.2 F 13.0  
8:00 222.7 F 8.3  
10:00 236.0 F 11.8  
12:00 240.1 F 12.1  
14:00 241.1 F 11.4  
16:00 255.0 F 18.6  
18:00 266.7 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.5 F 29.7  
2:00 271.5 F 29.2  
3:00 272.5 F 29.9  
4:00 271.9 F 29.2  
  
Min 271.4 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.7 F 0.0  
5:00 194.3 F 0.0  
10:00 202.2 F 0.0  
15:00 205.0 F 0.0  
20:00 203.1 F 0.0  
25:00 200.9 F 0.0  
30:00 199.8 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:27  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:16  
CYCLE COMPLETE

04/28/2022 02:00 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_