

-----  
Midmark M11 - v1.0.5  
Total Cycles: 586

Name: RITTER2  
Sterilizer ID: V2313228  
11/30/2021 09:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 106.5 F 0.0  
2:00 107.3 F 0.3  
4:00 118.4 F 1.9  
6:00 150.0 F 5.2  
8:00 178.1 F 9.1  
10:00 193.3 F 9.9  
12:00 221.0 F 6.7  
14:00 234.7 F 11.0  
16:00 229.2 F 7.0  
18:00 241.9 F 12.8  
20:00 240.7 F 11.1  
22:00 252.1 F 17.2  
24:00 262.9 F 23.6  
26:00 272.9 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.3  
1:00 272.4 F 30.1  
2:00 272.4 F 29.3  
3:00 272.3 F 30.1  
4:00 272.8 F 29.6

Min 271.5 F 29.1  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.0 F 0.0  
5:00 230.8 F 0.0  
10:00 209.7 F 0.0  
15:00 216.4 F 0.0  
20:00 206.7 F 0.0  
25:00 211.3 F 0.0  
30:00 211.3 F 0.0

Drying Complete

FILLING: 1:11  
HEATING: 26:12  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:04:28  
CYCLE COMPLETE

11/30/2021 10:51 AM  
ACCEPTED: X  
REJECTED:  
Operator: U04