
Midmark M11 - v1.0.5
Total Cycles: 1308

Name: RITTER2
Sterilizer ID: V2313228
07/15/2022 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.4 F 0.0
2:00 185.0 F 1.8
4:00 203.8 F 6.6
6:00 218.1 F 12.3
8:00 226.9 F 8.3
10:00 239.2 F 12.2
12:00 240.8 F 12.1
14:00 240.9 F 11.2
16:00 253.9 F 17.9
18:00 265.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.4 F 29.5
2:00 271.7 F 29.3
3:00 272.3 F 29.5
4:00 271.5 F 29.1

Min 271.5 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 217.2 F 0.0
10:00 227.8 F 0.0
15:00 231.1 F 0.0
20:00 226.1 F 0.0
25:00 222.0 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

07/15/2022 03:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____