
Midmark M11 - v1.0.5
Total Cycles: 1594

Name: RITTER2
Sterilizer ID: V2313228
10/13/2022 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 185.4 F 3.3
4:00 193.9 F 8.3
6:00 213.2 F 10.0
8:00 226.0 F 9.7
10:00 235.0 F 8.9
12:00 242.4 F 12.7
14:00 243.6 F 13.0
16:00 256.4 F 19.9
18:00 268.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.6
1:00 272.7 F 30.3
2:00 272.0 F 29.6
3:00 272.1 F 30.1
4:00 272.8 F 30.2

Min 271.6 F 29.4
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 201.2 F 0.0
10:00 237.9 F 0.0
15:00 221.7 F 0.0
20:00 218.6 F 0.0
25:00 216.7 F 0.0
30:00 230.3 F 0.0

Drying Complete

FILLING: 2:41
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

10/13/2022 12:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____