
Midmark M11 - v1.0.0
Total Cycles: 4719

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2024 04:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.9 F 0.0
2:00 127.5 F 1.7
4:00 151.5 F 4.5
6:00 174.1 F 7.7
8:00 197.3 F 11.9
10:00 208.7 F 3.1
12:00 220.3 F 6.6
14:00 230.5 F 10.3
16:00 233.0 F 8.5
18:00 239.5 F 11.0
20:00 235.8 F 8.5
22:00 248.6 F 14.5
24:00 259.2 F 20.3
26:00 269.3 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 29.0
2:00 272.1 F 28.4
3:00 271.6 F 28.4
4:00 272.7 F 29.1

Min 271.4 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 206.3 F 0.0
10:00 189.9 F 0.0
15:00 188.0 F 0.0
20:00 196.5 F 0.0
25:00 212.2 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:10
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:05:58
CYCLE COMPLETE

07/11/2024 05:48 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____