
Midmark M11 - v1.0.0
Total Cycles: 3593

Name: RITTER 1
Sterilizer ID: V2178198
07/31/2023 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 72.8 F 0.0
4:00 89.1 F 0.0
6:00 120.5 F 0.0
8:00 151.4 F 0.0
10:00 174.9 F 1.7
12:00 193.0 F 4.8
14:00 203.2 F 7.9
16:00 213.7 F 11.2
18:00 221.8 F 4.7
20:00 229.5 F 8.2
22:00 239.2 F 12.3
24:00 236.5 F 9.2
26:00 241.0 F 10.4
28:00 246.3 F 13.1
30:00 257.6 F 19.1
32:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.1 F 28.2
2:00 272.6 F 28.6
3:00 271.8 F 28.3
4:00 271.9 F 28.0

Min 271.4 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.9 F 0.0
5:00 181.3 F 0.0
10:00 188.4 F 0.0
15:00 190.7 F 0.0
20:00 189.0 F 0.0
25:00 191.0 F 0.0
30:00 209.9 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:21
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:11:28
CYCLE COMPLETE

07/31/2023 10:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____