
Midmark M11 - v1.0.5
Total Cycles: 1772

Name: RITTER2
Sterilizer ID: V2313228
11/23/2022 03:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.1 F 0.0
2:00 202.7 F 4.6
4:00 214.5 F 10.7
6:00 226.8 F 7.7
8:00 237.7 F 11.0
10:00 244.3 F 13.4
12:00 245.3 F 13.3
14:00 259.4 F 20.9
16:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.4
2:00 271.6 F 29.2
3:00 272.5 F 29.5
4:00 271.7 F 28.9

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 228.1 F 0.0
10:00 239.7 F 0.0
15:00 233.6 F 0.0
20:00 229.0 F 0.0
25:00 228.7 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 2:44
HEATING: 16:36
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:56:29
CYCLE COMPLETE

11/23/2022 04:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____