
Midmark M11 - v1.0.5
Total Cycles: 3997

Name: RITTER2
Sterilizer ID: V2313228
10/03/2024 04:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.3 F 0.0
2:00 150.2 F 2.5
4:00 166.3 F 6.0
6:00 190.1 F 10.6
8:00 216.9 F 4.8
10:00 227.7 F 9.4
12:00 235.3 F 10.2
14:00 237.2 F 12.0
16:00 238.5 F 10.6
18:00 251.4 F 16.9
20:00 262.5 F 23.5
22:00 272.5 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.7
1:00 272.4 F 30.1
2:00 271.6 F 30.1
3:00 271.7 F 29.7
4:00 271.8 F 30.1

Min 271.3 F 29.6
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 217.0 F 0.0
10:00 220.4 F 0.0
15:00 218.8 F 0.0
20:00 206.5 F 0.0
25:00 189.5 F 0.0
30:00 183.7 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

10/03/2024 05:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____