
Midmark M11 - v1.0.0
Total Cycles: 2412

Name: RITTER 1
Sterilizer ID: V2178198
10/06/2022 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.8 F 0.0
2:00 180.7 F 2.4
4:00 191.3 F 7.3
6:00 213.1 F 13.1
8:00 228.3 F 7.9
10:00 240.2 F 13.2
12:00 239.6 F 10.8
14:00 237.0 F 9.1
16:00 251.8 F 16.2
18:00 263.8 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.3 F 28.7
2:00 272.0 F 28.8
3:00 272.0 F 28.6
4:00 272.0 F 28.9

Min 271.5 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 242.5 F 0.0
10:00 231.2 F 0.0
15:00 217.3 F 0.0
20:00 220.8 F 0.0
25:00 239.5 F 0.0
30:00 226.5 F 0.0

Drying Complete

FILLING: 2:14
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

10/06/2022 03:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____