
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001796
10:24 19/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:54
STERILIZING 4:09
136.3 C 222kPa 4:09
136.3 C 224kPa 4:39
136.3 C 223kPa 5:09
136.3 C 223kPa 5:39
136.3 C 224kPa 6:09
136.4 C 225kPa 6:39
136.2 C 222kPa 7:09
136.4 C 225kPa 7:39
136.4 C 225kPa 7:39
VENTING 7:39
AIR DRYING 8:29
CYCLE COMPLETE 8:51

Digital Signature #
B8E8C496602412DF
