
Midmark M11 - v1.0.0
Total Cycles: 1910

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2022 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.8 F 0.0
2:00 184.5 F 4.0
4:00 196.7 F 8.7
6:00 214.2 F 9.8
8:00 224.5 F 8.0
10:00 239.1 F 13.2
12:00 237.9 F 10.2
14:00 237.4 F 8.1
16:00 249.5 F 15.0
18:00 261.4 F 21.7
20:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 28.9
2:00 271.4 F 28.4
3:00 272.5 F 28.9
4:00 271.4 F 28.3

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 209.0 F 0.0
10:00 210.3 F 0.0
15:00 221.5 F 0.0
20:00 233.8 F 0.0
25:00 229.4 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

05/05/2022 07:02 AM
ACCEPTED: X
REJECTED:
Operator: U02