

-----  
Midmark M11 - v1.0.5  
Total Cycles: 258

Name: RITTER2  
Sterilizer ID: V2313228  
07/28/2021 10:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 195.3 F 4.0  
4:00 220.1 F 9.6  
6:00 223.5 F 7.4  
8:00 241.9 F 12.5  
10:00 240.7 F 10.8  
12:00 238.9 F 10.0  
14:00 252.9 F 16.8  
16:00 264.6 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.5  
1:00 272.6 F 29.0  
2:00 271.9 F 29.0  
3:00 271.8 F 29.2  
4:00 272.0 F 29.4  
  
Min 270.8 F 28.3  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 226.2 F 0.0  
10:00 223.4 F 0.0  
15:00 226.3 F 0.0  
20:00 229.7 F 0.0  
25:00 231.1 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 2:25  
HEATING: 17:46  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:22  
CYCLE COMPLETE

07/28/2021 11:57 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01