
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000732
13:14 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:18
STERILIZING 3:17
136.4 C 226kPa 3:17
136.1 C 231kPa 3:47
136.2 C 232kPa 4:17
136.2 C 233kPa 4:47
136.4 C 233kPa 5:17
136.4 C 233kPa 5:47
136.4 C 233kPa 6:17
136.4 C 231kPa 6:47
136.4 C 231kPa 6:47
VENTING 6:47
AIR DRYING 7:39
CYCLE COMPLETE 8:10

Digital Signature #
7B65748D6EE72578
