
Midmark M11 - v1.0.0
Total Cycles: 2312

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2022 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 178.3 F 4.2
4:00 192.9 F 8.9
6:00 214.2 F 8.4
8:00 227.5 F 8.8
10:00 235.6 F 10.5
12:00 240.1 F 11.8
14:00 238.9 F 10.2
16:00 252.8 F 17.1
18:00 264.9 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.0 F 28.8
2:00 272.4 F 29.3
3:00 271.4 F 28.5
4:00 272.2 F 29.0

Min 271.4 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 221.4 F 0.0
10:00 240.7 F 0.0
15:00 237.6 F 0.0
20:00 225.9 F 0.0
25:00 223.2 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

09/12/2022 03:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____