
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 003046
16:15 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:28
STERILIZING 4:22
136.4 C 225kPa 4:22
136.4 C 222kPa 4:52
136.4 C 225kPa 5:22
136.4 C 221kPa 5:52
136.2 C 222kPa 6:22
136.3 C 222kPa 6:52
136.4 C 223kPa 7:22
136.3 C 224kPa 7:52
136.3 C 224kPa 7:52
VENTING 7:52
AIR DRYING 8:48
CYCLE COMPLETE 8:54

Digital Signature #
37CFF0E563131B2E
