
Midmark M11 - v1.0.5
Total Cycles: 2144

Name: RITTER2
Sterilizer ID: V2313228
03/14/2023 07:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.0 F 0.0
2:00 79.6 F 0.0
4:00 94.4 F 0.0
6:00 109.8 F 1.4
8:00 146.3 F 4.4
10:00 170.3 F 7.9
12:00 190.6 F 12.0
14:00 218.5 F 5.2
16:00 227.4 F 9.0
18:00 237.3 F 13.2
20:00 234.5 F 9.4
22:00 238.6 F 10.4
24:00 243.7 F 12.7
26:00 254.9 F 18.6
28:00 265.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.6 F 30.1
2:00 272.2 F 29.5
3:00 271.3 F 29.3
4:00 272.5 F 30.1

Min 271.2 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 214.7 F 0.0
10:00 232.3 F 0.0
15:00 230.4 F 0.0
20:00 228.7 F 0.0
25:00 229.5 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:60
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:02
CYCLE COMPLETE

03/14/2023 09:05 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____