
Midmark M11 - v1.0.0
Total Cycles: 4119

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2023 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.9 F 0.0
2:00 174.7 F 3.5
4:00 204.9 F 8.1
6:00 224.3 F 8.8
8:00 231.2 F 8.7
10:00 239.7 F 11.2
12:00 241.5 F 11.3
14:00 239.0 F 9.7
16:00 253.0 F 16.5
18:00 264.3 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.5 F 28.5
2:00 271.1 F 27.8
3:00 272.7 F 28.6
4:00 272.3 F 28.4

Min 271.0 F 27.7
Max 273.0 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 224.4 F 0.0
10:00 226.1 F 0.0
15:00 228.9 F 0.0
20:00 230.4 F 0.0
25:00 226.4 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

12/14/2023 04:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____