
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002175
13:03 28/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:41
STERILIZING 3:41
136.0 C 220kPa 3:41
136.3 C 224kPa 4:11
136.0 C 222kPa 4:41
136.2 C 224kPa 5:11
136.3 C 225kPa 5:41
136.4 C 226kPa 6:11
136.4 C 225kPa 6:41
136.4 C 225kPa 7:11
136.4 C 225kPa 7:11
VENTING 7:11
AIR DRYING 8:01
CYCLE COMPLETE 8:16

Digital Signature #
2479A331EB4F8D01
