

-----  
Midmark M11 - v1.0.0  
Total Cycles: 754

Name: RITTER 1  
Sterilizer ID: V2178198  
03/29/2021 10:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 120.7 F 0.0  
2:00 130.3 F 0.5  
4:00 136.6 F 2.8  
6:00 167.0 F 6.3  
8:00 189.9 F 11.0  
10:00 215.3 F 4.4  
12:00 224.2 F 9.5  
14:00 237.2 F 9.1  
16:00 241.8 F 11.9  
18:00 238.9 F 9.9  
20:00 252.7 F 16.5  
22:00 264.3 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.2  
1:00 272.7 F 28.8  
2:00 272.1 F 28.6  
3:00 271.9 F 28.8  
4:00 272.0 F 29.1

Min 270.9 F 28.1  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.5 F 0.0  
5:00 238.4 F 0.0  
10:00 233.1 F 0.0  
15:00 229.0 F 0.0  
20:00 223.4 F 0.0  
25:00 210.3 F 0.0  
30:00 202.4 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 23:51  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:01:43  
CYCLE COMPLETE

03/29/2021 11:13 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01