
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002499
11:38 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 2:59
STERILIZING 4:08
136.0 C 220kPa 4:08
136.0 C 223kPa 4:38
136.1 C 223kPa 5:08
136.2 C 225kPa 5:38
136.4 C 225kPa 6:08
136.4 C 224kPa 6:38
136.4 C 225kPa 7:08
136.4 C 225kPa 7:38
136.4 C 225kPa 7:38
VENTING 7:38
AIR DRYING 8:33
CYCLE COMPLETE 9:35

Digital Signature #
BFE56092AF076F68
