
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002201
14:40 01/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:55
STERILIZING 3:05
136.5 C 227kPa 3:05
136.4 C 226kPa 3:35
136.0 C 222kPa 4:05
136.3 C 225kPa 4:35
136.4 C 226kPa 5:05
136.4 C 226kPa 5:35
136.4 C 226kPa 6:05
136.3 C 225kPa 6:35
136.3 C 225kPa 6:35
VENTING 6:35
AIR DRYING 7:26
CYCLE COMPLETE 7:35

Digital Signature #
554712FC3787A9D1
