
Midmark M11 - v1.0.0
Total Cycles: 2615

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2022 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.6 F 0.0
2:00 209.8 F 4.6
4:00 209.4 F 10.3
6:00 221.9 F 5.5
8:00 235.7 F 11.7
10:00 236.4 F 9.5
12:00 237.9 F 8.7
14:00 250.0 F 15.4
16:00 262.5 F 22.7
18:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.6 F 29.2
2:00 272.0 F 29.0
3:00 272.7 F 29.2
4:00 271.8 F 28.6

Min 271.7 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.7 F 0.0
5:00 224.1 F 0.0
10:00 233.6 F 0.0
15:00 226.2 F 0.0
20:00 226.2 F 0.0
25:00 228.4 F 0.0
30:00 229.9 F 0.0

Drying Complete

FILLING: 3:19
HEATING: 18:10
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

11/28/2022 04:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____