
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000762
16:49 02/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:43
STERILIZING 4:58
136.4 C 223kPa 4:58
136.1 C 219kPa 5:28
136.3 C 225kPa 5:58
136.3 C 224kPa 6:28
136.4 C 223kPa 6:58
136.1 C 221kPa 7:28
136.4 C 224kPa 7:58
136.3 C 223kPa 8:28
136.3 C 223kPa 8:28
VENTING 8:28
AIR DRYING 9:23
CYCLE COMPLETE 9:24

Digital Signature #
E84DBD7FCD6654D6
