
Midmark M11 - v1.0.5
Total Cycles: 2495

Name: RITTER2
Sterilizer ID: V2313228
06/09/2023 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.5 F 0.0
2:00 148.1 F 0.1
4:00 168.3 F 3.2
6:00 188.6 F 7.3
8:00 207.5 F 11.8
10:00 219.5 F 5.9
12:00 232.0 F 10.4
14:00 233.3 F 8.4
16:00 243.7 F 13.2
18:00 243.0 F 12.1
20:00 255.1 F 18.4
22:00 265.9 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.8
2:00 272.1 F 29.2
3:00 271.2 F 28.8
4:00 272.1 F 29.7

Min 271.2 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 219.5 F 0.0
10:00 218.0 F 0.0
15:00 197.5 F 0.0
20:00 195.6 F 0.0
25:00 201.5 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:49
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

06/09/2023 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____