
Midmark M11 - v1.0.5
Total Cycles: 3336

Name: RITTER2
Sterilizer ID: V2313228
01/30/2024 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.0 F 0.0
2:00 143.2 F 0.8
4:00 153.4 F 3.5
6:00 181.3 F 7.7
8:00 203.1 F 12.6
10:00 219.9 F 6.6
12:00 233.4 F 11.7
14:00 233.5 F 9.0
16:00 239.3 F 10.1
18:00 246.9 F 14.1
20:00 259.1 F 20.8
22:00 270.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 271.6 F 28.8
2:00 271.8 F 29.0
3:00 272.0 F 29.3
4:00 272.2 F 29.6

Min 271.4 F 28.8
Max 272.2 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 221.3 F 0.0
10:00 222.4 F 0.0
15:00 224.1 F 0.0
20:00 230.9 F 0.0
25:00 225.4 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

01/30/2024 12:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____