
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001347
9:50 15/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:36
PRESSURIZING 4:36
STERILIZING 5:50
136.0 C 220kPa 5:50
135.9 C 221kPa 6:20
136.3 C 224kPa 6:50
136.1 C 222kPa 7:20
136.3 C 222kPa 7:50
136.4 C 225kPa 8:20
136.4 C 225kPa 8:50
136.3 C 224kPa 9:20
136.3 C 224kPa 9:20
VENTING 9:20
AIR DRYING 10:18
CYCLE COMPLETE 18:51

Digital Signature #
DB71D7F880360D76
