
Midmark M11 - v1.0.5
Total Cycles: 821

Name: RITTER2
Sterilizer ID: V2313228
02/22/2022 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.0 F 0.0
2:00 171.5 F 2.6
4:00 189.0 F 7.9
6:00 211.4 F 8.8
8:00 228.9 F 9.4
10:00 235.6 F 9.2
12:00 241.3 F 11.1
14:00 248.8 F 15.0
16:00 262.1 F 22.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 271.9 F 28.9
2:00 271.8 F 28.9
3:00 272.0 F 29.0
4:00 272.1 F 29.2

Min 271.6 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 222.3 F 0.0
10:00 226.6 F 0.0
15:00 228.4 F 0.0
20:00 229.1 F 0.0
25:00 229.3 F 0.0
30:00 228.0 F 0.0

Drying Complete

FILLING: 2:08
HEATING: 17:60
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:15
CYCLE COMPLETE

02/22/2022 12:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____