
Midmark M11 - v1.0.0
Total Cycles: 3847

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2023 10:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.3 F 0.0
2:00 192.7 F 2.9
4:00 198.9 F 8.0
6:00 217.7 F 13.4
8:00 226.2 F 8.7
10:00 236.0 F 10.9
12:00 241.4 F 11.6
14:00 239.6 F 10.2
16:00 253.3 F 16.9
18:00 265.0 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.4 F 28.7
2:00 271.5 F 28.3
3:00 272.1 F 28.3
4:00 272.1 F 28.8

Min 271.4 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 216.0 F 0.0
10:00 213.4 F 0.0
15:00 234.3 F 0.0
20:00 231.3 F 0.0
25:00 228.8 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:42
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

10/02/2023 11:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____