
Midmark M11 - v1.0.0
Total Cycles: 639

Name: MIDMARK
Sterilizer ID: V2178198
01/28/2021 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.2 F 0.0
2:00 124.6 F 0.5
4:00 142.3 F 2.5
6:00 172.6 F 6.2
8:00 197.0 F 11.2
10:00 220.1 F 4.9
12:00 229.0 F 10.2
14:00 231.7 F 7.4
16:00 244.0 F 12.7
18:00 241.0 F 10.9
20:00 254.7 F 17.5
22:00 266.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.0 F 28.4
2:00 272.3 F 28.9
3:00 271.5 F 28.0
4:00 272.3 F 28.7

Min 271.3 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 225.3 F 0.0
10:00 238.7 F 0.0
15:00 227.1 F 0.0
20:00 226.5 F 0.0
25:00 227.6 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:23
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

01/28/2021 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____