
Midmark M11 - v1.0.5
Total Cycles: 3462

Name: RITTER2
Sterilizer ID: V2313228
03/20/2024 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.7 F 0.0
2:00 94.3 F 0.0
4:00 107.6 F 0.2
6:00 141.8 F 2.6
8:00 163.7 F 5.7
10:00 183.8 F 9.3
12:00 201.6 F 13.1
14:00 212.8 F 6.7
16:00 225.2 F 11.0
18:00 230.3 F 7.6
20:00 242.7 F 13.4
22:00 242.7 F 12.3
24:00 254.5 F 18.5
26:00 265.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 271.8 F 30.1
2:00 272.4 F 29.9
3:00 272.6 F 30.4
4:00 271.9 F 29.8

Min 270.6 F 29.0
Max 273.3 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 226.0 F 0.0
10:00 220.7 F 0.0
15:00 214.7 F 0.0
20:00 195.2 F 0.0
25:00 187.9 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 27:45
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:06:22
CYCLE COMPLETE

03/20/2024 04:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____