
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006134
10:41 16/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:03
STERILIZING 3:06
136.5 C 222kPa 3:06
136.2 C 220kPa 3:36
136.4 C 225kPa 4:06
136.1 C 221kPa 4:36
136.0 C 219kPa 5:06
136.3 C 223kPa 5:36
136.4 C 224kPa 6:06
136.4 C 224kPa 6:36
136.4 C 224kPa 6:36
VENTING 6:36
AIR DRYING 7:26
CYCLE COMPLETE 9:15

Digital Signature #
5AF81550A5BD9BB5
