
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000449
9:40 01/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:46
STERILIZING 3:55
136.0 C 225kPa 3:55
136.2 C 232kPa 4:25
136.0 C 230kPa 4:55
136.1 C 232kPa 5:25
136.0 C 231kPa 5:55
136.2 C 231kPa 6:25
136.4 C 233kPa 6:55
136.4 C 233kPa 7:25
136.4 C 233kPa 7:25
VENTING 7:25
AIR DRYING 8:21
CYCLE COMPLETE 8:22

Digital Signature #
AAE4BD0543F57DCA
