
Midmark M11 - v1.0.5
Total Cycles: 2659

Name: RITTER2
Sterilizer ID: V2313228
07/18/2023 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 77.5 F 0.0
4:00 108.2 F 0.0
6:00 105.9 F 0.0
8:00 144.5 F 1.8
10:00 168.0 F 5.1
12:00 186.7 F 8.6
14:00 202.7 F 12.4
16:00 216.0 F 5.8
18:00 226.4 F 9.9
20:00 234.0 F 9.7
22:00 239.1 F 11.5
24:00 237.1 F 9.6
26:00 250.0 F 15.9
28:00 261.0 F 22.2
30:00 271.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.6 F 29.8
2:00 271.4 F 29.2
3:00 272.5 F 29.9
4:00 271.6 F 29.1

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 209.0 F 0.0
10:00 208.8 F 0.0
15:00 204.9 F 0.0
20:00 202.9 F 0.0
25:00 203.3 F 0.0
30:00 202.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

07/18/2023 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____