

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1257

Name: RITTER 1  
Sterilizer ID: V2178198  
10/12/2021 02:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.5 F 0.0  
2:00 192.0 F 2.6  
4:00 215.4 F 7.2  
6:00 229.5 F 12.5  
8:00 228.9 F 7.5  
10:00 238.9 F 12.5  
12:00 237.1 F 9.8  
14:00 239.4 F 8.9  
16:00 248.4 F 14.5  
18:00 260.0 F 20.9  
20:00 270.4 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 272.5 F 29.2  
2:00 272.0 F 28.5  
3:00 272.0 F 28.8  
4:00 272.5 F 29.0  
  
Min 271.4 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 226.2 F 0.0  
10:00 216.8 F 0.0  
15:00 200.5 F 0.0  
20:00 196.7 F 0.0  
25:00 209.1 F 0.0  
30:00 209.1 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 20:44  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

10/12/2021 03:21 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01