
Midmark M11 - v1.0.0
Total Cycles: 5091

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2024 12:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.2 F 0.0
2:00 172.4 F 0.4
4:00 177.3 F 1.1
6:00 200.2 F 4.5
8:00 210.5 F 8.8
10:00 224.7 F 13.3
12:00 230.4 F 8.7
14:00 241.9 F 11.4
16:00 242.9 F 12.1
18:00 241.3 F 10.9
20:00 254.5 F 17.4
22:00 265.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.4 F 28.8
2:00 271.6 F 28.1
3:00 271.9 F 28.7
4:00 272.0 F 28.5

Min 271.1 F 27.9
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 208.7 F 0.0
10:00 202.4 F 0.0
15:00 210.9 F 0.0
20:00 228.0 F 0.0
25:00 229.0 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:02:30
CYCLE COMPLETE

11/14/2024 01:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____