
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000366
7:25 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:27
STERILIZING 3:32
136.2 C 229kPa 3:32
136.1 C 231kPa 4:02
136.1 C 230kPa 4:32
136.0 C 230kPa 5:02
136.2 C 231kPa 5:32
136.3 C 232kPa 6:02
136.4 C 233kPa 6:32
136.4 C 233kPa 7:02
136.4 C 233kPa 7:02
VENTING 7:02
AIR DRYING 7:54
CYCLE COMPLETE 9:53

Digital Signature #
38ED3F6434D19171
