

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1027

Name: RITTER 1  
Sterilizer ID: V2178198  
07/19/2021 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.6 F 0.0  
2:00 165.1 F 1.2  
4:00 164.2 F 3.8  
6:00 186.8 F 7.2  
8:00 211.2 F 11.7  
10:00 223.6 F 6.2  
12:00 229.0 F 11.1  
14:00 231.7 F 7.6  
16:00 245.0 F 13.5  
18:00 243.1 F 12.0  
20:00 255.2 F 18.2  
22:00 266.1 F 24.8  
24:00 273.1 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 271.8 F 29.4  
2:00 272.4 F 29.4  
3:00 272.6 F 29.4  
4:00 272.7 F 29.3

Min 271.5 F 29.0  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 235.3 F 0.0  
10:00 227.8 F 0.0  
15:00 227.8 F 0.0  
20:00 222.9 F 0.0  
25:00 207.9 F 0.0  
30:00 160.1 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 24:13  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:02:36  
CYCLE COMPLETE

07/19/2021 10:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01