
Midmark M11 - v1.0.5
Total Cycles: 49

Name: RITTER2
Sterilizer ID: V2313228
05/27/2021 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.2 F 0.0
2:00 162.1 F 2.5
4:00 177.6 F 5.7
6:00 193.7 F 9.7
8:00 212.8 F 9.5
10:00 223.9 F 7.8
12:00 235.8 F 12.5
14:00 236.8 F 10.1
16:00 238.8 F 9.9
18:00 248.8 F 14.9
20:00 259.7 F 20.9
22:00 269.5 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.5
2:00 271.9 F 28.9
3:00 271.4 F 28.8
4:00 272.3 F 29.5

Min 271.3 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 203.0 F 0.0
10:00 197.8 F 0.0
15:00 192.4 F 0.0
20:00 188.4 F 0.0
25:00 186.1 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 23:06
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:26
CYCLE COMPLETE

05/27/2021 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____