
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001777
16:18 15/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:04
STERILIZING 3:04
136.0 C 220kPa 3:04
136.4 C 223kPa 3:34
136.0 C 222kPa 4:04
136.4 C 221kPa 4:34
136.0 C 222kPa 5:04
136.4 C 225kPa 5:34
136.3 C 224kPa 6:04
136.4 C 225kPa 6:34
136.4 C 225kPa 6:34
VENTING 6:34
AIR DRYING 7:24
CYCLE COMPLETE 7:55

Digital Signature #
E2DB329A4AC7D14A
