
Midmark M11 - v1.0.5
Total Cycles: 2916

Name: RITTER2
Sterilizer ID: V2313228
09/25/2023 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.6 F 0.0
2:00 89.8 F 0.0
4:00 75.4 F 0.0
6:00 100.7 F 0.0
8:00 141.0 F 0.0
10:00 168.2 F 0.7
12:00 187.1 F 3.5
14:00 204.1 F 6.6
16:00 221.4 F 9.8
18:00 231.7 F 13.4
20:00 231.9 F 7.9
22:00 241.8 F 12.1
24:00 236.7 F 9.2
26:00 241.1 F 10.5
28:00 246.2 F 13.3
30:00 256.9 F 19.1
32:00 266.8 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.3 F 29.0
2:00 271.5 F 28.7
3:00 272.5 F 28.9
4:00 271.4 F 28.7

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 247.0 F 0.0
10:00 211.7 F 0.0
15:00 178.4 F 0.0
20:00 174.2 F 0.0
25:00 172.5 F 0.0
30:00 171.9 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:37
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:11:51
CYCLE COMPLETE

09/25/2023 10:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____