

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001820  
8:15 10/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:54  
PRESSURIZING 2:00  
STERILIZING 3:02  
136.5 C 225kPa 3:02  
136.4 C 225kPa 3:32  
136.4 C 226kPa 4:02  
136.3 C 224kPa 4:32  
136.2 C 221kPa 5:02  
136.2 C 224kPa 5:32  
136.3 C 224kPa 6:02  
136.4 C 226kPa 6:32  
136.4 C 226kPa 6:32  
VENTING 6:32  
AIR DRYING 7:23  
CYCLE COMPLETE 13:46  
-----

Digital Signature #  
ACDBC31EDB17209F  
-----