
Midmark M11 - v1.0.0
Total Cycles: 717

Name: MIDMARK
Sterilizer ID: V2178198
03/02/2021 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.5 F 0.0
2:00 107.0 F 0.0
4:00 123.6 F 0.2
6:00 158.6 F 2.7
8:00 192.1 F 6.7
10:00 208.8 F 11.8
12:00 223.4 F 6.2
14:00 230.7 F 11.5
16:00 232.9 F 8.2
18:00 240.5 F 10.6
20:00 245.4 F 12.9
22:00 258.2 F 19.7
24:00 269.4 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.7 F 29.4
2:00 272.8 F 29.1
3:00 272.6 F 29.3
4:00 273.0 F 29.5

Min 270.9 F 28.3
Max 273.3 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.8 F 0.0
5:00 238.4 F 0.0
10:00 221.0 F 0.0
15:00 215.9 F 0.0
20:00 220.3 F 0.0
25:00 231.0 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:07
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:03:04
CYCLE COMPLETE

03/02/2021 11:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____